

# Integra®

## Tissue Technologies Product Portfolio

	Integra® Dermal Regeneration Template	Integra® Meshed Dermal Regeneration Template	Integra® Bilayer Wound Matrix	Integra® Meshed Bilayer Wound Matrix	Integra® Wound Matrix	Integra® Wound Matrix (Thin)	Integra® Flowable Wound Matrix	Integra® PriMatrix®	Integra® PriMatrix® Ag	
	Orange	Green	Teal / Blue	Green	Purple	Yellow	Blue	Purple	Silver	
<b>Composition</b>	A porous matrix of cross-linked bovine tendon collagen and glycosaminoglycan and a semi-permeable silicone layer	A porous matrix of cross-linked bovine tendon collagen and glycosaminoglycan and a semi-permeable silicone layer	A porous matrix of cross-linked bovine tendon collagen and glycosaminoglycan and a semi-permeable silicone layer	A porous matrix of cross-linked bovine tendon collagen and glycosaminoglycan and a semi-permeable silicone layer	A porous matrix of cross-linked bovine tendon collagen and glycosaminoglycan	A porous matrix of cross-linked bovine tendon collagen and glycosaminoglycan that has 50% less collagen compared to each of the corresponding sq cm sizes of Integra® Wound Matrix	Granulated cross-linked bovine tendon collagen and glycosaminoglycan	An acellular dermal tissue matrix derived from fetal bovine dermis	An acellular dermal tissue matrix derived from fetal bovine dermis with Ionic Silver	
<b>Indications</b>	<p>Postexcisional treatment of life-threatening full-thickness or deep partial-thickness thermal injuries where sufficient autograft is not available at the time of excision or not desirable due to the physiological condition of the patient.</p> <p>The repair of scar contractures when other therapies have failed or when donor sites for repair are not sufficient or desirable due to the physiological condition of the patient.</p>	<p>Postexcisional treatment of life-threatening full-thickness or deep partial-thickness thermal injuries where sufficient autograft is not available at the time of excision or not desirable due to the physiological condition of the patient.</p> <p>The repair of scar contractures when other therapies have failed or when donor sites for repair are not sufficient or desirable due to the physiological condition of the patient.</p>	<p>The management of wounds including: partial and full-thickness wounds, pressure ulcers, venous ulcers, diabetic ulcers, chronic vascular ulcers, surgical wounds (donor sites/grafts, post-Moh's surgery, post-laser surgery, podiatric, wound dehiscence), trauma wounds (abrasions, lacerations, second-degree burns, and skin tears) and draining wounds</p>	<p>The management of wounds including: partial and full-thickness wounds, pressure ulcers, venous ulcers, diabetic ulcers, chronic vascular ulcers, surgical wounds (donor sites/grafts, post-Moh's surgery, post-laser surgery, podiatric, wound dehiscence), trauma wounds (abrasions, lacerations, second-degree burns, and skin tears) and draining wounds. Integra® Meshed Bilayer Wound Matrix may be used in conjunction with negative pressure wound therapy</p>	<p>The management of wounds including: partial and full-thickness wounds, pressure ulcers, venous ulcers, diabetic ulcers, chronic vascular ulcers, tunneled/undermined wounds, surgical wounds (donor sites/grafts, post-Moh's surgery, post-laser surgery, podiatric, wound dehiscence), trauma wounds (abrasions, lacerations, second-degree burns, and skin tears) and draining wounds</p>	<p>The management of wounds including: partial and full-thickness wounds, pressure ulcers, venous ulcers, diabetic ulcers, chronic vascular ulcers, tunneled/undermined wounds, surgical wounds (donor sites/grafts, post-Moh's surgery, post-laser surgery, podiatric, wound dehiscence), trauma wounds (abrasions, lacerations, second-degree burns, and skin tears) and draining wounds</p>	<p>The management of wounds including: partial and full-thickness wounds, pressure ulcers, venous ulcers, diabetic ulcers, chronic vascular ulcers, tunneled/undermined wounds, surgical wounds (donor sites/grafts, post-Moh's surgery, post-laser surgery, podiatric, wound dehiscence), trauma wounds (abrasions, lacerations, second-degree burns, and skin tears) and draining wounds</p>	<p>The management of wounds that include: partial and full thickness wounds, pressure ulcers, diabetic ulcers, venous ulcers, second-degree burns, surgical wounds (donor sites/grafts, post-Mohs surgery, post-laser surgery, podiatric, wound dehiscence), trauma wounds (abrasions, lacerations, and skin tears), tunneled/undermind wounds and draining wounds</p>	<p>The management of wounds that include: partial and full thickness wounds, pressure ulcers, diabetic ulcers, venous ulcers, second-degree burns, surgical wounds (donor sites/grafts, post-Mohs surgery, post-laser surgery, podiatric, wound dehiscence), trauma wounds (abrasions, lacerations, and skin tears), tunneled/undermind wounds and draining wounds</p>	
<b>Contraindications</b>	This device should not be used in patients with known sensitivity to bovine collagen or chondroitin materials	This device should not be used in patients with known sensitivity to bovine collagen or chondroitin materials	<p>This device should not be used in patients with known sensitivity to bovine collagen or chondroitin materials.</p> <p>This device is not indicated for third degree burns</p>	<p>This device should not be used in patients with known sensitivity to bovine collagen or chondroitin materials.</p> <p>This device is not indicated for third degree burns</p>	<p>This device should not be used in patients with known sensitivity to bovine collagen or chondroitin materials.</p> <p>This device is not indicated for third degree burns</p>	<p>This device should not be used in patients with known sensitivity to bovine collagen or chondroitin materials.</p> <p>This device is not indicated for third degree burns</p>	<p>This device should not be used in patients with known sensitivity to bovine collagen or chondroitin materials.</p> <p>This device is not indicated for third degree burns</p>	<p>PriMatrix® is not designed, sold, or intended for use except as indicated. PriMatrix® should not be used for patients with a known history to collagen or bovine products.</p> <p>This device is not indicated for use in third-degree burns.</p>	<p>PriMatrix® is not designed, sold, or intended for use except as indicated. PriMatrix® should not be used for patients with a known history to collagen or bovine products.</p> <p>This device is not indicated for use in third-degree burns.</p>	
<b>Sizes</b>	5cm x 5cm (2in x 2in) 10cm x 12.5cm (4in x 5in) 10cm x 25cm (4in x 10in) 20cm x 25cm (8in x 10in)	5cm x 5cm (2in x 2in) 10cm x 12.5cm (4in x 5in) 10cm x 25cm (4in x 10in) 20cm x 25cm (8in x 10in)	5cm x 5cm (2in x 2in) 10cm x 12.5cm (4in x 5in) 10cm x 25cm (4in x 10in) 20cm x 25cm (8in x 10in)	5cm x 5cm (2in x 2in) 10cm x 12.5cm (4in x 5in) 10cm x 25cm (4in x 10in) 20cm x 25cm (8in x 10in)	5cm x 5cm (2in x 2in) 10cm x 12.5cm (4in x 5in) 10cm x 25cm (4in x 10in) 20cm x 25cm (8in x 10in)	5cm x 5cm (2in x 2in) 10cm x 12.5cm (4in x 5in) 10cm x 25cm (4in x 10in) 20cm x 25cm (8in x 10in)	5cm x 5cm (2in x 2in) 10cm x 12.5cm (4in x 5in) 10cm x 25cm (4in x 10in)	3cc	<p>Solid:</p> <p>8cm x 8cm 8cm x 12cm 10cm x 12cm 10cm x 25cm</p> <p>Fenestrated:</p> <p>8cm x 8cm 8cm x 12cm 10cm x 25cm 20cm x 25cm</p> <p>Meshed:</p> <p>8cm x 8cm 8cm x 12cm 10cm x 25cm</p>	<p>Solid:</p> <p>8cm x 8cm 8cm x 12cm 10cm x 12cm 10cm x 25cm</p> <p>Fenestrated:</p> <p>8cm x 8cm 10cm x 25cm 20cm x 25cm</p> <p>Meshed:</p> <p>10cm x 25cm</p>
<b>Box</b>										

For more information or to place an order, please contact:

United States, Canada, Asia, Pacific, Latin America  
USA 877-444-1122 • 866-800-7742 fax  
Outside USA 609-936-5400 • 609-750-4259 fax  
[integralife.com](http://integralife.com)

PriMatrix, NeuraGen, TenoGlide, BioFix, Integra and the Integra logo are registered trademarks of Integra LifeSciences Corporation or its subsidiaries in the United States and/or other countries. NeuraWrap is a trademark of Integra LifeSciences Corporation.  
©2016 Integra LifeSciences Corporation. All rights reserved. Printed in USA. oM 0441308-2-EN

**INTEGRA**  
LIMIT UNCERTAINTY